



Drum

HEARING INSTRUMENT TEST BOX (HIT)

DESCRIPTION

Drum is a hearing instrument test box (HIT), with a high ambient noise attenuation level. Drum gives you the chance to perform all the tests aimed at hearing aids analysis via air conduction.

STANDARDS

EN 60118-7/ ANSI S3.22

AVAILABLE MEASUREMENT MODALITIES

Single-curve or multi-curves
Standard or custom measurements

AVAILABLE TESTS

- OSPL90
- Full On Gain (FOG)
- I/O
- Attack/Release
- Frequency Response
- Equivalent Input Noise
- THD
- Battery Drain
- Telecoil
- Frequency Measurement (sweep or broadband)

AVAILABLE OUTPUTS

Loudspeaker or Telecoil


AVAILABLE STIMULI

Pure tone, sweep or broadband stimuli

HIT: FREQUENCY, INTENSITY AND MEASUREMENTS

Frequency range: 125 Hz – 10 kHz
Frequency resolution: 1/3, 1/6, 1/12, 1/24 per octave
Intensity range: 40 – 100 dB SPL
Coupler intensity range: 40 – 145 dB
Reference microphone measurement: Omni-Directional
Telecoil stimulation range: 10-100 mA/m

ATTENUATION FEATURES



* WN @ 90 dB RMS

1kHz attenuation: 33dB SPL

INTERNAL CHAMBER

Egg-shaped (Minimum sound reflection)
Internal dimensions: 15x25x13cm
H.A. position: Floating, on acoustically-transparent material, in front of the loudspeaker
Loudspeaker features: 3", full range

AVAILABLE COUPLERS

2cc coupler for hearing aids:

- ITE
- RIC
- BTE

BATTERY SIMULATORS

Standard 5A, 10A, 312, 13, 675

MONITORING

Operator interface for coupler monitoring

CALIBRATION

Validity: 12 months

COMPUTER INTERFACE

USB connection (driverless)
Compatible software: Inventis Maestro (Noah compatible)

PC MINIMUM REQUIREMENTS

Processor: Intel® i3 or above
Display: 1280x1024
Operative system: Windows 7 or above, 32 or 64 bits

POWER SUPPLY

Power supply: 15V, 2A cont., through an external medical grade 100-240 Vac 50/60 Hz power supply

MECHANICS

Size (WxDxH): 32 x 33 x 22 cm / 12.6 x 13 x 8.7 in
Weight: 7.5 kg / 16.5 lbs

APPLICABLE STANDARDS

HIT: EN 60118-7 / ANSI S3.22
Electrical safety: EN 61010-1
EMC: EN 61326-1

CE CERTIFICATE

93/42/EEC classification: Class I
Classification rule (Annex IX, 93/42/EEC): 12

INCLUDED PARTS

- Reference microphone
- Coupler microphone
- 2cc cavity
- Couplers for ITE, BTE, RIC
- Putty for h.a. adapter
- BTE coupler tubes
- Clips for BTE h.a. positioning
- Coupler foam support
- Battery simulator
- Inventis Software Suite
- User manual
- Medical power supply
- USB cable